

Certificate of Analysis

**Product: ARTSMedia Oocyte Aspiration Medium, 20 ml
(AM Oocyte Aspiration)**

Catalog No. 502020

Lot No.:	AYYASPBXX
Expiry date:	YYYY-MM-DD
HSA Lot No.:	

Quality Control	Test Method	Specification	Results
pH (37°C)	Ph Eur. 2.2.3, USP <791>	7.2 - 7.45	
Osmolality	Ph Eur. 2.2.35, USP <785>	270-300 mOsm/kg	
Sterility	Ph Eur. 2.6.1, USP<71>	No growth	
Mouse Embryo Assay	FDA guidance "Mouse Embryo Assay for Assisted Reproduction Technology Devices, January" 2021	1-Cell Mouse Embryo Assay (MEA) ≥80% of embryos developed to expanded blastocyst at 96 hours.	
Endotoxin	USP <85>	< 0.25 EU/ml	
HSA Concentration	HPLC	2-3 g/l	
Appearance	Visual inspection	Clear and particulate-free	

Release date and signature:

QC Manager

Version 1.0, 2024JAN22

Rx only

Company address:

ARTSMedia Denmark ApS
Kongevejen 149
2830 Virum
Denmark

Company details:

CRN: 41808845
VAT number: DK41808845
Phone: +45 53 50 45 30
info@artsmedia.dk

Bank account:

Bank: Sparekassen Sjælland
Registration number: 6070
Account number: 0000379518
SWIFT: SWESDK22
IBAN: DK3560700000379518