

Certificate of Analysis

Product AM-IVC Medium

Product No.: 501010

Following batches included:

Lot Number: A24IVCBo1

Expiry date: 2025/09/20

HSA lot no.: K331A687A


10 ml

Quality Control		Specifications	Results
pH (37°C)	Ph Eur. 2.2.3, USP <791>	7.2 - 7.45	7.34
Osmolality	Ph Eur. 2.2.35, USP <785>	257 - 275 mOsm/kg	265 mOsm/kg
Sterility: Ph Eur. 2.6.1, USP<71>		No growth	No growth
Endotoxin: USP <85>		<0.05 EU/ml	<0.05 EU/ml
Appearance visual inspection		Clear and particulate free	Passed
HSA Concentration (HPLC)		2-3 mg/ml	2.3 mg/ml
Mouse Embryo Assay Test *)		1-cell to expanded blastocyst within 96 hours: ≥ 80 %	Passed

*) FDA guidance "Mouse Embryo Assay for Assisted Reproduction Technology Devices, January" 2021

Product released for sale in compliance with ARTSMedia Denmark quality standards.

Date 2024/04/19 and Signature:


Christina Kjempff, Quality Control Manager
Lotte Stroebech, CEO

Rx only

Version 6.0, 2024FEB08

Company address:

ARTSMedia Denmark ApS
Kongevejen 149
2830 Virum
Denmark

Company details:

CRN: 41808845
VAT number: DK41808845
Phone: +45 53 50 45 30
info@artsmedia.dk

Bank account:

Bank: Sparekassen Sjælland
Registration number: 6070
Account number: 0000379518
SWIFT: SWESDK22
IBAN: DK3560700000379518