

Certificate of Analysis

AM-IVC Medium

ARTSMedia In Vitro Culture Medium

Product No.: 501010

Following batches included:

Lot Number: A25IVCB01

Expiry date: 2026/08/20

HSA lot no.: M424A6871


10 ml

Quality Control		Specifications	Results
pH (37°C)	Ph Eur. 2.2.3, USP <791>	7.2 - 7.45	7.38
Osmolality	Ph Eur. 2.2.35, USP <785>	257 - 275 mOsm/kg	262 mOsm/kg
Sterility	Ph Eur. 2.6.1, USP<71>	No growth	No growth
Endotoxin	USP <85>	<0.05 EU/ml	<0.05 EU/ml
Appearance visual inspection		Clear and particulate free	Passed
HSA concentration (HPLC)		2-3 mg/ml	2.4 mg/ml
Mouse Embryo Assay Test		1-cell to expanded blastocyst within 96 hours: ≥ 80 %	Passed

Product released for sale in compliance with ARTSMedia Denmark quality standards.

Date 2025/03/17 and Signature:


Christina Kjempff, Quality Control Manager


Lotte Stroebech, CEO

Rx only

Version 9.0, 2024OCT23

Company address:

ARTSMedia Denmark ApS
Kongevejen 149
2830 Virum
Denmark

Company details:

CRN: 41808845
VAT number: DK41808845
Phone: +45 53 50 45 30
info@artsmedia.dk

Bank account:

Bank: Sparekassen Sjælland
Registration number: 6070
Account number: 0000379518
SWIFT: SWESDK22
IBAN: DK3560700000379518