

Certificate of Analysis

AM-SDG Medium 45 %

ARTSMedia Sperm Density Gradient Medium 45 %

Product No.: 503100

Following batches included:

Lot Number: AYYGRAUB29

Expiry date: 2026/06/05

☐ 20 ml

☐ 50 ml

☒ 100 ml

Quality Control

pH Ph Eur. 2.2.3, USP <791>
Osmolality Ph Eur. 2.2.35, USP <785>
Sterility, Ph Eur. 2.6.1, USP<71>: :
Endotoxin: USP <85>
HSSA (human Sperm Survival Test)

Specifications

7.20-7. 90
310-340 mOsm/kg
No growth
<0.5 EU/ml
% motility compared with
control after 24 hours: $\geq 75\%$
1.0500-1.0700 g/ml


Results

7.26
323 mOsm/kg
No growth
<0.5 EU/ml
Passed
1.0560 g/ml

Product released for sale in compliance with ARTSMedia Denmark quality standards.

Date 2025/02/05 and Signature:


Christina Kjempff, Quality Control Manager


Lotte Stroebech, CEO

Rx only

Version 3.0, 2024OCT23

Company address:

ARTSMedia Denmark ApS
Kongevejen 149
2830 Virum
Denmark

Company details:

CRN: 41808845
VAT number: DK41808845
Phone: +45 53 50 45 30
info@artsmedia.dk

Bank account:

Bank: Sparekassen Sjælland
Registration number: 6070
Account number: 0000379518
SWIFT: SWESDK22
IBAN: DK356070000379518